Dated: 12/07/2024

NATIONAL INSTITUTE OF TECHNOLOGY KARNATAKA, SURATHKAL

DEPARTMENT OF MECHANICAL ENGINEERING POST SRINIVASNAGAR, MANGALORE – 575 025 (D K)

Phone: (0824) 2474000. Fax: (0824) 2474033

E- mail: info@nitk.ac.in Website: http://www.nitk.ac.in



NOTICE INVITING QUOTATION

Notification No: NITK/MECH/DPD/OH35/IRG/2024-25/PKK/FHWS/04

Name of Goods	Fume Hood with Sink
Estimated Amount	Rs.1,20,000/-
EMD Amount (2% of estimated amount) (Applicable for estimate above 2.5 lakhs)	NA
Time for Supply of item after release of Purchase order	14 Days
Document Download / Sale Start Date	15/07/2024, at 09:00 AM
Clarification Start Date	15/07/2024, at 09:00 AM
Clarification End Date	22/07/2024, at 05:00 AM
Bid Submission Start Date	15/07/2024, at 09:00 AM
Last Date for submission of bids	30/07/2024, before 3.00 PM
Bid Opening Date	31/07/2024, at 09:00 AM
Address for Submission of bids [Type Address of the Department and Contact Person Details with Phone Number]	Dr. Poornesh Kumar Koorata, Associate Professor, Department of Mechanical Engineering, NITK, Srinivasnagar, Mangaluru - 575025 Contact No.: +91-824-247-3650 Email ID: kpkoorata@nitk.edu.in



NATIONAL INSTITUTE OF TECHNOLOGY KARNATAKA, SURATHKAL DEPARTMENT OF MECHANICAL ENGINEERING

POST SRINIVASNAGAR, MANGALORE - 575 025

Phone: (0824) 2474000 Fax: (0824) 2474033 E- mail: info@nitk.ac.in Website: http://www.nitk.ac.in

Notification No: NITK/MECH/DPD/OH35/IRG/2024-25/PKK/FHWS/04

NOTICE INVITING QUOTATION (NIQ)

The National Institute of Technology Karnataka, Surathkal (in short – NITK, Surathkal) is an Institute Of National Importance Under Ministry of Education Govt of India, imparting Technical Education and engaged in Research Activities. It is proposed to procure the items for the departmental academic/research activities.

Sealed Quotations as per the Price Schedule given in this NIQ are invited for the following items subject to the terms and conditions, from the reputed manufacturers or its authorised dealers so as to reach on or before scheduled date and time. The quotations in the firm's Business letter head should be address to the "Director, NITK, Surathkal". The envelope shall be superscribed with the Quotation Notification Number and the Name of the Goods for which quotation is submitted.

Name of Goods	Fume Hood with Sink
Estimated Amount	Rs.1,20,000/-
EMD Amount (2% of estimated amount) (Applicable for estimate above 2.5 lakhs)	NA
Time for Supply of item after release of Purchase order	14 Days
Document Download / Sale Start Date	15/07/2024, at 09:00 AM
Clarification Start Date	15/07/2024, at 09:00 AM
Clarification End Date	22/07/2024, at 05:00 AM
Bid Submission Start Date	15/07/2024, at 09:00 AM
Last Date for submission of bids	30/07/2024, before 3.00 PM
Bid Opening Date	31/07/2024, at 09:00 AM

Sd/[Signature of HOD with Seal]

Note: Institute shall not be responsible for any postal delay about non-receipt /non-delivery of the bids or due to wrong addressee.

Date: 12/07/2024

SECTION-1 Terms and Conditions

- 1. The rates should be quoted for preferably FOR destination from supply within India.
- 2. The bidder shall indicate the excise duty exemption for the goods if applicable.
- 3. The rate quoted should be on unit basis. Taxes and other charges should be quoted separately, considering exemptions if any. The rate should be quoted in INR only
- 4. Rate quoted should be inclusive of Testing, commissioning and Installation of equipment and Training.
- 5. Payment: No advance payment will be made. Payment will be made only after the supply of the item in good and satisfactory condition and receipt of performance security by supplier.
- 6. Guarantee/Warrantee period should be specified for the complete period should be specified in section 3 of this tender document.
- 7. Period requirement for the supply and installation of item should be specified in section 3 of this tender document.
- 8. In case of dispute, the matter will be subject to Mangalore Jurisdiction only.
- 9. <u>Delivery Period</u>: The Delivery Period/Time shall be deemed to be the essence of the Contract and delivery must be completed not later than such date(s). If the Supplier does not perform its obligations within the Delivery Period/Date mentioned in the Contract, the same would constitute the breach of the Contract and the Institute shall have the right to cancel or withdraw the Contract for the unsupplied portion after the expiry of the original or extended delivery date or period stipulated in the Contract. Such cancellation of the contract on account of non performance by the Supplier would entitle the Buyer to forfeit the EMD/Performance Security besides other actions such as debarment from the Institute as per GFR 2017.
- 10. <u>Liquidated Damages</u>: Timely delivery is the essence of the contract and hence if the Supplier fails to deliver Goods within the original/extended delivery period(s) specified in the contract, the Institute will be entitled to deduct/recover the Liquidated Damages for the delay, unless covered under Force Majeure conditions aforesaid, @ 1% per week or part of the week of the delayed period as pre-estimated damages not exceeding 5% of the contract value without any controversy/dispute of any sort whatsoever.
- 11. The bidder needs to comply with restrictions under rule No 144 (xi) of GFR (General Financial Rules) 2017 and the clause pertaining to Public Procurement Division Order (Public Procurement no 1,2 & Division Order (Public Procurement no 1,

- 12. The bidder needs to submit a Self-certification in line with Make in India, Public Procurement Order No. P-45021/2/2017-BE-II dt. 15.06.2017, P 45021/2/2017-PP (BE-II) dated 28.05.2018, P-45021/2/2017-PP (BE-II) dated 29.05.2019 and P45021/2/2017- PP (BE II) dated 16.09.2020 of DPIIT, Ministry of Commerce & Industry. The Bidder who will not meet this criterion will be rejected in the technical bid. The Self-certification should be provided as per Annexure C.
- 13. EMD must be in the form of Bank Guarantee Annexure-D (Fixed Deposit Receipt/Insurance surety bonds/DD/e-Payment mode (through Credit/Debit cards / Net Banking). It will be valid for 180 days from the date of opening of the tender. In case of EMD payment made through E-Payment, UTR Number with details should be uploaded. The firm registered with MSME/NSIC as per MSME procurement policy order 2012 will be exempted from submission of EMD. Intended parties will have to give proof of registration along with their tender. EMD of the unsuccessful bidders shall be refunded without any interest at the earliest after finalization of the purchase of the concerned item.
 - a. The EMD will be returned to the BIDDERs(s) whose offer is not accepted by NITK within one month from the date of the placing of the purchase order(s) on the selected BIDDER(s). In the case of the BIDDER(s) whose offer is accepted the EMD will be returned on submission of Performance Security in the form of Bank Guarantee (BG). However, if the return of EMD is delayed for any reason, no interest/penalty shall be payable to the Bidders.
 - b. The successful BIDDER, on the award of contract/order, must send the contract/order acceptance in writing, within 15 days of award of contract/order failing which the EMD will be forfeited.
 - c. The EMD shall be forfeited in case a successful BIDDER fails to furnish the Performance Security.
 - d. Bidder should prepare the EMD as per the instructions specified in the tender document. The original EMD should be posted/couriered/given in person to the concerned official, latest by the last date of bid submission as specified in the tender documents. The details of the BG/DD/FD/any other accepted instrument, physically sent, should tally with the details available in the scanned copy and the data entered during bid submission time. Otherwise, the uploaded bid will be rejected.
- 14. <u>Performance Security</u>: The successful bidder shall deposit performance security of 3% of the quoted price in the form of Bank Guarantee (ANNEXURE-E) / Demand Draft / Fixed Deposit Receipt of any scheduled bank drawn in favour of The Director, NITK Surathkal, Mangaluru payable at Mangaluru. In case the bidder fails to deposit the said Performance Security within the stipulated period, including the extended period if any, the Earnest Money Deposited by the bidder shall be forfeited automatically without any notice. Please note the following points:
 - a. Successful bidder should submit performance Security as prescribed above to the purchase Section,
 NITK Surathkal, on or before 30 days from the date of issue of order acknowledgment.
 - b. Performance Security shall be for the due and faithful performance of the contract and shall remain binding, notwithstanding such variations, alterations for extensions of time as may be made, given, conceded, or agreed to between the successful bidder and the purchaser under the terms and conditions of acceptance to tender.
 - c. The successful bidder is entirely responsible for the due performance of the contract in all respects according to the terms and conditions of the tender.
 - d. The validity of the Performance Security must cover the warranty period plus two months.

SCHEDULE OF REQUIREMENTS, SPECIFICATIONS AND ALLIED DETAILS

[To be filled up by the Department / Centre of NITK, Surathkal]

: NA

Item(s) Name to be Procured : Fume Hood with Sink

Brief Specifications of the Item(s)

(Attach Additional Sheet if necessary)

: Specification Sheet is attached

Quantity : 01

Any other details / requirement : NA

Warranty Period required : 01 Year

Delivery Schedule expected

after placement of Purchase order

(in Weeks) : 02 Weeks

Performance Security to be given

by Successful Bidder after release of

Purchase Order (in Rupees)

(applicable for estimate Above 2.5 lakhs)

SECTION 3 PRICE SCHEDULE

[To be used by the bidder for submission of the quotation]

1.	Item Name	:			
2.	Specifications (Conforming to Schedule of requirements Enclose additional sheets if necessary)	:			
3.	Currency and Unit Price	:			
4.	Quantity	:			
5.	Item Cost (SI No. 3 * SI. No. 4)	:			
6.	Taxes and Other Charges (i) Specify the type of taxes and duties in percentages and also in figures. (ii) Specify Other Charges in figures.	:			
7.	Warranty Period (Conforming to the Schedule of requirements)	:			
8.	Delivery Schedule (Conforming to the Schedule of requirements)	:			
9.	Name and address of the Firm for placing purchase order	:			
10.	Name and address of Indian authorized agent (in case of imports only)	:			
Signat	ure of the Bidder:				
Name	and Designation:			-	
Busine	ess Address :				
				-	
Place:		;	Seal of the	Bidder's Fir	m

SECTION 4 CONTRACT FORM

[To be provided by the bidder in the business letter head]

- 1. (Name of the Supplier's Firm) hereby abide by the delivery schedule mentioned in this document for supply of the items if the purchase order is awarded.
- 2. The item will be supplied conforming to the specifications stated in this document without any defect and deviations.
- 3. Warranty will be given for the period mentioned in this document and service will be rendered to the satisfaction of NITK, Surathkal during this period.

Signature of the Bidd	er:	
Name	:	
Business Address	:	
Place : Date :		Seal of the Bidder's Firm

Annexure-A

(Compliance to be submitted in the bidder's letterhead) (Applicable for estimate above 2.5 lakhs)

Sub: Compliance to Government of India order OM No.6/18/2019-PPD dated 23.07.2020 & 24.7.2020 regarding restrictions under Rule 144 (XI) of the General Financial Rules (GFRs), 2017

Item Name:	
Enquiry No.:	
pertaining to the procurement no	(name of the bidder company) have read the clauses Department of Expenditure's (DoE) Public Procurement Division Order (Public 1,2 & 3 vide ref. F.No.6/18/2019-PPD dated 23.07.2020 & 24.7.2020) regarding rocurement from a bidder of a country that shares a land border with India.
We hereby certitender.	fy that we are not from such a country and eligible to be considered for this
,	pliance of above said Gol Order and its subsequent amendment, (if any), by any ad to commercial rejection of their bids by NITK)
For and behalf of	f(Name of the bidder)
(Signature, date	& seal of an authorized representative of the bidder)

Annexure-B

(Compliance to be submitted in the bidder's letterhead) (Applicable for estimate above 2.5 lakhs)

Sub: Compliance to Government of India order OM No.6/18/2019-PPD dated 23.07.2020 & 24.7.2020 regarding restrictions under Rule 144 (XI) of the General Financial Rules (GFRs), 2017

Item Name:	
Enquiry No.:	
We M/s	(name of the bidder company) have read the clauses
	epartment of Expenditure's (DoE) Public Procurement Division Order (Public 2 & 3 vide ref. F.No.6/18/2019-PPD dated 23.07.2020 & 24.7.2020) regarding
	curement from a bidder of a country which shares a land border with India.
We are from such	a country which shares a land border with India & have been registered with
the Competent Aut	hority as specified in the above-said order. We hereby certify that we fulfill all s regard and are eligible to be considered.
Evidence of valid re	egistration by the Competent Authority is attached.
	ance of above said Gol Order and its subsequent amendment, (if any), by any
bidder(s) shall lead	to commercial rejection of their bids by NITK)
For and babalf of	
For and benait of_	(Name of the bidder)
(Signature, date &	seal of an authorized representative of the bidder)
, •	•

Annexure-C

Self-Certification on the letterhead of the company (Applicable for estimate above 5 lakhs)

In line with Government Public Procurement Order No. P-45021/2/2017-BE-II dt. 15.06.2017, P-45021/2/2017-PP (BE-II) dated 28.05.2018, P-45021/2/2017-PP (BE-II) dated 29.05.2019 and P-45021/2/2017-PP (BE-II) dated 16.09.2020, we hereby certify that (Supplier name) are local supplier meeting requirement of minimum local content % defined in as above orders for the material against Enquiry / Tender No Details of the location at which local value addition will be made are as follows: We also understand false declarations will be in breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151 (iii) of the General Financial Rules along with such other actions as may be permitted under law. Date: Signature: Place: Name and Designation: Mobile no: Office Telephone No: Email ID:

Office Seal:

FORMAT FOR EARNEST MONEY DEPOSIT / BID BOND

(Applicable for estimate above 2.5 lakhs)

(To be typed on Non-judicial stamp paper of value Indian Rupees One Hundred) (TO BE ESTABLISHED THROUGH ANY OF THE NATIONAL BANKS (WHETHER SITUATED AT MANGALURU OR OUTSTATION) WITH A CLAUSE TO ENFORCE THE SAME ON THEIR LOCAL BRANCH AT MANGALURU OR ANY SCHEDULED BANK (OTHER THAN NATIONALISED BANK) SITUATED AT MANGALURU. BONDS ISSUED BY COOPERATIVE BANKS ARE NOT ACCEPTED)

LETTER OF GUARANTEE To
Director, National Institute of Technology Karnataka, Srinivasnagar P.O., Surathkal Mangaluru – 575025
IN ACCORDANCE WITH YOUR TENDER for the supply of
supply
During the validity of this Bank Guarantee: We,
reservation, protest, demur and recourse. Any such demand made by theNITK Surathkal shall be conclusive and recourse. Any such demand made by the purchaser shall be binding on the Bank irrespective of any dispute or difference raised by the Bidder. The Guarantee shall be irrevocable and shall remain valid up to(180 days from the date of issue of Bank Guarantee) If any further extension is required, the same shall be extended to such required period on receiving instruction from the Bidder, on whose behalf the is Guarantee is issued.
Notwithstanding anything contained herein:

* Our liability under this Bank Guarantee shall not exceed Rs...... (Rupees......).

- * This Bank Guarantee shall be valid up to.....(date).
- * We are liable to pay the guaranteed amount or any part thereof under this Bank Guarantee Only and only if you serve upon us a written claim on or before(date).

Yours truly,

Signature and seal of the guarantor: Name of Bank:

Address:

Date:

Instruction to Bank: Bank should note that on expiry of Bond Period, the Original Bond will not be returned to the Bank. The bank is requested to take appropriate necessary action on or after the expiry of the bond period

Annexure-E

FORMAT FOR PERFORMANCE GUARANTEE BOND

(To be typed on Non-judicial stamp paper of the value of Indian Rupees of One Hundred)
(Applicable for estimate above 2.5 lakhs)

To, Director, National Institute of Technology Karnataka, Srinivasnagar P.O., Surathkal Mangaluru – 575025
Whereas
And Whereas you have stipulated it in the said contract that the contractor shall furnish you with a bank guarantee by a Commercial bank for the sum specified therein as security for compliance with its obligations as per the contract;
And Whereas we have agreed to give the contractor such a bank guarantee.
Now Therefore we hereby affirm that we are guarantors and responsible to you, on behalf of the contractor, up to a total of
We hereby waive the necessity of your demanding the sail debt from the contractor before presenting us with demand.
We further agree that no change or addition to or other modification of the terms of the contract to be performed thereunder or of any of the contract documents which may be made between you and the contractor shall in any way release us from any liability under this guarantee, and we hereby waive notice of any such change, addition, or modification.
This guarantee shall be valid until theday of20
Our*(Name & Address of the*(Name & Address of the*(branch) is liable to pay the guaranteed amount depending on the filing of a claim and any part thereof under this Bank Guarantee only and only if you serve upon us at our* branch a written claim or demand and received by us at our* branch on or before Dt otherwise, the bank shall be discharged of all liabilities under this guarantee after that.

(Signature of the authorized officer of the Bank)	
Name and designation of the officer	
Seal, name & address of the Bank and address of Branch	

*Preferably at the headquarters of the authority competent to sanction the expenditure for the procurement of goods or at the concerned district headquarters or the state headquarters.

Specifications of Fume hood with Sink

Laboratory Fume Hoods used for removing the exhaust toxic and harmful vapours protecting them from inhalation of harmful vapours in Chemical laboratory. So, the fume hood should have been designed per international standards and passed international tests for its quality.

Description	Technical Specifications
Fumehood Details	
Type of Fumehood	Benchtop type fumehood
Main Body fume hood	MS powder
Inside Cupboard	FRP lining (should be alkali and acid resistant)
Working table $(L \times B)$	$4 \text{ feet} \times 2 \text{ feet}$
Required base material of the inner work region	Granite
Number of Drawers	03 drawer
Fume Exhaust System Sp	ecifications
Internal Washing or Cleaning	Yes
Storage Cabinet Depth	750 mm
Casing and Coupling Thickness	06 mm
Exhaust motor assembly power	2.0 HP
Shutters and Drawer Fronts Thickness	9 mm
Storage Cabinet Width	1600 mm
Ducting Diameter	300 mm
Internal Thickness	10 mm
Ducting Material	FRP
Fume Depth	800 mm

Storage Cabinet Height	850 mm
Fume Width	1600 mm
Blower exhaust system	Blower cover with FRP
Exhaust Motor Assembly	200 V
Voltage	
Exhaust Motor Assembly	2000 rpm
Speed	
Lighting	LED lights of 20 watts with acidic fume resistant casing
Suction Rate	>/= 1000 cfm
Hod top thickness	15 mm
Sash Handle	Anodized Aluminium Handle
Sash Type	Single vertical sliding sash
External Thickness	15 mm
Base Cabinet Features	Adjustable Shelf
Ducting Thickness	6 mm
Material Construction	Mild Steel
Fume Height	1500 mm
Storage Cabinet	17 mm
Thickness	17 111111
Fume Hood Workshop	20 mm thick polypropylene material with skirting of on 3 sides
(Bed)	20 mm thek polypropytene material with skirting of on 3 sides
Casing and Coupling	FRP Bisphenol
Material	TRI Displicitor
Exported Face Velocity at	100 rpm
open sash	100 Ipin
Baffle thickness	28 mm
Exhaust Motor Assembly	Three Phase
Phase	Timee Timee
Scrubber Specifications	
Mist Eliminator	PP Mist Eliminator
Packing	50 mm PP Pall rings
Pump for Water	1 HP single-phase centrifugal pump with PP Impeller and casing
recirculation	1 111 single phase centifican pump with 11 impenci and casing

Glass Water Level Indicator Water Drain Outlet Diameter with PP ball valve Vapour Outlet Diameter Water Circulation Flow Rate Scrubber Height Scrubber Type Scrubber Flow rate Construction of Scrubber Material Yes, the indicator must be provided along with the equipment 25 mm 40 LPM Cylindrical type spray tower scrubber 2900 mm 2 mm thick PP + 3 mm thick FRP
Indicator Water Drain Outlet Diameter with PP ball valve Vapour Outlet Diameter Water Circulation Flow Rate Scrubber Height Scrubber Type Cylindrical type spray tower scrubber Scrubber Flow rate Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
Diameter with PP ball valve Vapour Outlet Diameter 200 mm Water Circulation Flow Rate Scrubber Height 2900 mm Scrubber Type Cylindrical type spray tower scrubber Scrubber Flow rate 4000 cfm Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
valve Vapour Outlet Diameter 200 mm Water Circulation Flow Rate 40 LPM Scrubber Height 2900 mm Scrubber Type Cylindrical type spray tower scrubber Scrubber Flow rate 4000 cfm Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
Vapour Outlet Diameter 200 mm Water Circulation Flow Rate 40 LPM Scrubber Height 2900 mm Scrubber Type Cylindrical type spray tower scrubber Scrubber Flow rate 4000 cfm Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
Water Circulation Flow Rate 40 LPM Scrubber Height 2900 mm Scrubber Type Cylindrical type spray tower scrubber Scrubber Flow rate 4000 cfm Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
Rate Scrubber Height 2900 mm Scrubber Type Cylindrical type spray tower scrubber Scrubber Flow rate 4000 cfm Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
Rate Scrubber Height 2900 mm Scrubber Type Cylindrical type spray tower scrubber Scrubber Flow rate 4000 cfm Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
Scrubber Type Cylindrical type spray tower scrubber Scrubber Flow rate 4000 cfm Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
Scrubber Flow rate 4000 cfm Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
Construction of Scrubber 2 mm thick PP + 3 mm thick FRP
2 mm thick PP + 3 mm thick FRP
Scrubber Diameter 1300 mm
Recirculation Water Tank 250 litres
Capacity
Liquid Inlet Diameter 25 mm
Installation Details
On-site Installation The product should be delivered, installed, commissioned as
demonstrated for satisfactory performance as per technic
specifications at laboratory site.
Certification ISO 9001:2015 certified; CE certified (TCF/178)
Warranty Details
Worrenty 01 Year Comprehensive Warranty with on-site service. Further
Warranty service requests /complaints must be attended within 24 hours.
Internal Sink Specification
Sink and Tap (Total 01 Polypropylene internal sink (10" Diameter) of high quality wi
Quantity) single-way water tap (PVC) should be provided
Additional Requirement
Retort Stand Compatible Retort stand – Polypropylene (01 Quantity) with clam
must be provided with the Fume hood

Gas supply accessories	Compatible Brass Nozzle, Gas tube and knob controller must be provided
Exhaust Pipeline requirements	Compatible PVC pipe, as per the lab location, must be supplied with necessary fixtures (cowl cover, clamps, nuts, bolts, screws and others) for expelling the fume gases outside the lab.
Water supply accessories	Water pipeline connection with knob, tap and drain out tube (Polypropylene) must be provided.
Electrical Arrangement	The hood super structures should be fully wired (good grade). It should also have 01 set of electrical sockets and switches of (230V, 15/5A, 50Hz) of reputed make. MCBs with blower switch carrying built in starter and light switches in front side.
Cable Entering port	Cable and wires should be fire retardant grade. Port is needed for easy access of cable of the fume hood to the electrical socket.
Cabinets	Bottom Mounted Cabinets with horizontal shelves and stainless-steel door handle
Colour	Preferably WHITE / BLUE