



**NATIONAL INSTITUTE OF TECHNOLOGY KARNATAKA
SURATHKAL, SRINIVASNAGAR, MANGALORE-575025**

Date: 05.02.2025

Corrigendum No. 01

This is with reference to the tender notification for BOD Incubator, **Tender Reference No: NITK/CHE/IRG/NABL/24-25/BOD/1074 Dated: 07.01.2025**, There are following corrections:

- 1. Last Date and Time for submission of tender is extended up to 03.00 p.m. on 10.02.2025**
- 2. Date of opening of technical bid: 10.02.2025 at 04.00pm**

Additions/corrections to Technical Specification:

1. Capacity reduced to 3001 (from 735L); dimensions: Internal - 56x45x120 cm, External - 71x76x188 cm.
2. Control system must include a digital display, PT100 sensor, and microcontroller-based PID controller with 500-line memory for recording, storing, and printing set/process values, date & time, and variable print intervals, with RS-485 communication for PC interface.
3. Certifications required: IQ, OQ, PQ documents, Material Test Certificates, Calibration Reports, and 21CFR-PART11 compliance for software.
4. Material: Corrosion-resistant preferred; 3" CFC-free PUF insulation mandatory, with a see-through acrylic door (with illumination) and a 50mm validation port.

Proof for above mentioned specifications must be attached to the tender doc

**Sd/-
Prof. P. E Jagadeesh Babu
(Buyer)**

**Sd/-
HoD
Department of Chemical Engineering**